

ABSTRACT:

A method of forming a magnetic resonance image involves separate measurement of the position of a measuring site. The magnetic resonance image is corrected on the basis of the measured position of the measuring site. [Notably the temperature is changed at the measuring site, for example for laser ablation of tissue.]

5

[Fig. 1]